



MINIMET

File saving problems in SkyeLynx softwares


If you are using the SkyeLynx Standard, SkyeLynx Auto or SkyeLynx Deluxe softwares then the following settings are advised for your workstation:

1. Log on to Administrator mode and check that the software works normally in this mode
2. If possible, always use your workstation in full Administrator mode

If this is not permissible then:

3. Allow full read / write capabilities to the following areas on the hard drive:
 - A) C:\root directory
 - B) C:\skyedata folder
 - C) C:\skyeback folder
 - D) C:\lynxauto folder
 - E) C:\lynxdelx folder

If the above settings are not used you will find that data downloaded from the DataHog or MiniMet datalogger cannot be saved onto the hard drive.



SKYE INSTRUMENTS LTD
21, Ddole Enterprise Park, Llandrindod Wells, Powys, LD1 6DF, UK
Tel: +44(0)1597 824811 Fax: +44(0)1597 824812
Email: skyemail@skyeinstruments.com Web: www.skyeinstruments.com

